




America

CERTIFICATE

No. U8 112484 0003 Rev. 00

Holder of Certificate: **PhotoDynamic Inc.**
 1344 Summer St.
 Halifax NS B3H 0A8
 CANADA

Certification Mark:



Product: **General Medical Devices**
Therapeutic LED Light source

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 7169007747-100

Date, 2022-09-29

(Peter Keith)



America

CERTIFICATE

No. U8 112484 0003 Rev. 00

Model(s): PhytoLight 1.0 (REF: PL1.0)

Brand Name: PHOTODYNAMIC

Brand:  PHOTODYNAMIC

Tested according to: ANSI/AAMI ES60601-1:2005/(R)2012
 CSA C22.2 No. 60601-1:2014
 CSA C22.2 No. 60601-1-6:2011/A1:2015-10
 CSA C22.2 No. 60601-1-11:2015
 CSA C22.2 No. 60601-2-57:2011

Exclusions: All excluding Biocompatibility (cl. 11.7) and EMC (cl. 17)

Parameters:

Rated input voltage: 5 Vdc, Internally Powered
 Rated input current: 1 A
 Degree of protection: Type BF applied part
 Duty cycle: 1 min ON, 2 min OFF
 Protection Class: Class II